

MIS Warranty: MIS exercises great care and effort in maintaining the superior quality of its products. All MIS products are guaranteed to be free from defects in material and workmanship. However, should a customer find any fault in any MIS product after using it according to the directions, the defective product will be replaced.



Overview.

The MIS EZ-Base system features a unique Ti-Base, designed with a highly accessible screw channel for extreme angulation, making the process of anterior or posterior restorations simpler and more convenient.

CN-EZB15
VN-EZB15
CS-EZB15
CW-EZB15

MN-EZB15
MD-EZB15
MW-EZB15

CN-LEZ15
VN-LEZ15
CS-LEZ15
CW-LEZ15

MT-ELR10
MT-ESR10

MT-ELM10
MT-ESM10

Advantages.

Comfort and Safety

Restoration placement has never been simpler than with the EZ-Base system. Extreme angulation is safely handled within the screw channel. More angle options means more comfort for the clinician performing anterior and posterior restorations with convenient handling and placement.

EZ Procedure and Reliability

The EZ-Base screwdriver features a unique tip, which allows safe and reliable access from multiple angles, as well as gripping, tightening and loosening within the angulated screw-channel at a torque and with the convenience that are similar to a straight screw-channel.

Esthetics

The EZ-Base system opens up a world of options for prosthetic restorations in the esthetic zone. Where as in the past screw-retained restorations may have been ruled out for many anterior zone cases, the EZ-Base system provides a solution.



Specialized Components.

Unique screw and driver tip design enable extreme angulation and a wide path of insertion.



MD-E0220
MN-E0160



Restoration Options.

1.

EZ-Base abutment

Required Components:

Scan Post
EZ-Base Abutment with specialized prosthetic screw
Long and Short EZ-Base Motor Key
Long and Short EZ-Base Ratchet Key

1. Scanning
2. Designing and manufacturing
3. Cementing
4. Adjusting
5. Securing the abutment and final restoration

1.



2.



3.

4.

5.

Scan post positioning: thin wall towards the angulation of screw-channel, except for V3 NP cases where the thin wall of scan post should be positioned to the buccal.

Restoration Options.

Required Components:

EZ-Base Abutment with specialized prosthetic screw
Long and Short EZ-Base Motor Key
Long and Short EZ-Base Ratchet Key

2.

Adjustable EZ-Base abutment
Conventional method

1. Implant level impression taking
2. Abutment adjustments
3. Pattern fabrication
4. Casting and porcelain firing
5. Cementing
6. Securing the abutment and final restoration


1.


2.


3.


4.


5.


6.

Restoration Options.

Required Components:

EZ-Base Abutment with specialized prosthetic screw
Long and Short EZ-Base Motor Key
Long and Short EZ-Base Ratchet Key

3.

Adjustable EZ-Base abutment
CAD/CAM method

1. Implant level impression taking
2. Abutment adjustments
3. Scanning
4. Design and Manufacturing
5. Cementing
6. Securing the abutment and final restoration

1.

2.



3.

4.



5.

6.





MIS Implants Technologies Ltd.

www.mis-implants.com

The MIS Quality System complies with the following international quality standards: ISO 13485:2016 - Quality Management System for Medical Devices and Medical Device Directive 93/42/EEC. Please note that not all products are registered or available in every country or region.